



Kenya Power

Central Office - P.O. Box 30099 - 00100
Telephone – 254 – 02 – 3201000
Stima Plaza, Kolobot Road
Nairobi, Kenya
www.kenyapower.co.ke

Our ref: KP1/9A/9AA.1A/OT/31/16-17

20th July, 2017

Your ref:

RE: ADDENDUM NO.1 TO THE TENDER NO. KP1/9A/9AA.1A/OT/31/16-17 FOR PURCHASE OF TOOLS & EQUIPMENT FOR CUSTOMER SERVICE.

Please refer to the above Tender.

We make the following clarifications and amendments to the principal Tender Document (hereinafter abbreviated as the PTD) for the supply of tools and equipments for Customer Service dated July 2017.

1. RELATIONSHIP WITH THE PRINCIPAL TENDER DOCUMENT.

Save where expressly amended by the terms of this Addendum, the PTD shall continue to be in full force and effect. The provisions of this addendum shall be deemed to have been incorporated in and shall be read and constructed as part of the PTD.

2. AMMENDMENT TO SECTION IV-SCHEDULE OF REQUIREMENTS.

Item No.	Item	Description	KPLC Response
1	Single phase meter test bench	specs, point 4.2.14, “The <i>test bench shall have 40 position, 20 on each side</i> ”. On the other hand, at Appendix B the drawing, it shows 10 positions. Also, given max dimensions, 400(W) x 205(H) feels like 20 pos bench. Clarify whether to adopt the stated specifications in 4.2.14 or the drawings requirements	Refer to clause – 4.2.14 – The test bench shall have 40 positions, 20 on each side Also Amended Appendix B

2	Portable Working Standard class 0.05	In your specification clauses 4.3.6, 4.3.7 & 4.3.12 is requesting for voltage & current but references standard meters is only related to power & energy. Can the product be offered with an option of power quality analyzer combined?	Clauses 4.3.6, 4.3.7 & 4.3.12 maintained and should be adhered to. Acceptable on condition that the product offered meets all requirements for the portable working standard class 0.05.
---	--------------------------------------	--	---

3. AMMENDMENT TO SECTION XII A-MANUFACTURER'S AUTHORIZATION FORM.

Item No.	Item	Description	KPLC Response
1	Manufacture's Authorization Form	One of the documents required to be submitted by tenderer is Manufacturer Authorization. Is it required against tenderer to submit SOLE authorization?	Provide document as proof confirming you are the manufacturer of the said goods.

4. CLOSING DATE.


The closing date has changed from 25th July, 2017 at 10:00 am to 18th August, 2017 at 10:00 am. All other terms and conditions remain as per the Principal Tender Document (PTD)

Yours faithfully,
For: THE KENYA POWER & LIGHTING CO. LTD.


BERNARD NGUGI
GENERAL MANAGER, SUPPLY CHAIN

KP1/9A/9AA.1A/OT/31/16-17 PURCHASE OF TOOLS & EQUIPMENT FOR CUSTOMER SERVICE.

APPENDIX B:_METER RACK AND ACCESSORIES



NOTE

The bench is 40 position.
It shall have 2 sides each consisting of two rows of testing positions.
Each side with 20 positions numbered as above

