

**DOCUMENT NO.: KP1/6C/13D/1/TSP/14/023**



**Kenya Power**

**PORTABLE THREE PHASE WORKING STANDARD FOR TESTING ELECTRICITY METERS -  
SPECIFICATION**

A Document of the Kenya Power & Lighting Co. Ltd  
September 2023





**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 1 of 18</b>  |                                |

**Table of Contents**

|      |                                                   |    |
|------|---------------------------------------------------|----|
| 0.1  | CIRCULATION LIST .....                            | 2  |
| 0.2  | AMENDMENT RECORD .....                            | 3  |
|      | FOREWORD.....                                     | 4  |
| 1.   | SCOPE.....                                        | 5  |
| 2.   | NORMATIVE REFERENCES.....                         | 5  |
| 3.   | DEFINITIONS AND ABBREVIATIONS.....                | 5  |
| 4.   | REQUIREMENTS.....                                 | 6  |
| 4.1. | SERVICE CONDITIONS .....                          | 6  |
| 4.2. | DESIGN AND CONSTRUCTION FEATURES.....             | 6  |
| 4.3. | RATING.....                                       | 8  |
| 4.4. | COMPUTER REQUIREMENTS.....                        | 9  |
| 5.   | ACCESSORIES .....                                 | 10 |
| 6.   | TESTS REQUIREMENTS .....                          | 10 |
| 7.   | MARKING, LABELLING AND PACKING .....              | 10 |
|      | APPENDICES.....                                   | 11 |
| A.   | TESTS AND INSPECTION (Normative) .....            | 11 |
| B.   | QUALITY MANAGEMENT SYSTEM (Normative) .....       | 12 |
| C.   | DOCUMENTATION AND DEMONSTRATION (Normative).....  | 12 |
| D.   | GUARANTEED TECHNICAL PARTICULARS (Normative)..... | 14 |

|                                                                                                    |                                                                                                     |
|----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                 |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 2 of 18</b>  |                                |

#### 0.1 CIRCULATION LIST


| COPY NO. | COPY HOLDER                                                                                                                                                |
|----------|------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1        | Manager, Standards                                                                                                                                         |
| 2        | Electronic copy (pdf) on Kenya Power server( <a href="http://172.16.1.40/dms/browse.php?fFolderId=23">http://172.16.1.40/dms/browse.php?fFolderId=23</a> ) |

#### REVISION OF KPLC STANDARDS

In order to keep abreast of progress in the industry, KPLC Standards shall be regularly reviewed. Suggestions for improvements to approved standards, addressed to the Manager, Standards Department, are welcome.

**© Kenya Power & Lighting Co. Plc**

Copyright, Users are reminded that by virtue of Section 25 of the Copyright Act, 2001 (Revised 2014) Cap 130 of the Laws of Kenya, Copyright subsists in all KPLC Standards and except as provided under Section 26 of this Act, no KPLC Standard(s) may be reproduced, stored in retrieval system by any means without prior permission from the Managing Director & CEO, KPLC.

|                                                                                                    |                                                                                                      |
|----------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                  |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                              |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 3 of 18</b>  |                                |

## 0.2 AMENDMENT RECORD

| Rev No.               | Date<br>(YYYY-MM-DD) | Description of Change                                                                                                                                                                                                                                          | Prepared by<br>(Name &<br>Signature) | Approved by<br>(Name &<br>Signature) |
|-----------------------|----------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------|--------------------------------------|
| Issue 3               | September 2009       |                                                                                                                                                                                                                                                                |                                      |                                      |
| Issue 4               | December 2014        |                                                                                                                                                                                                                                                                | Aggrey Machasio                      | Joshua Mutua                         |
| Issue 5               | 2022-06-07           | New Issue                                                                                                                                                                                                                                                      | Nancy Wairimu                        | Eng. Simon<br>Kimatei                |
| Issue 5<br>Revision 1 | 2023-09-28           | 1.Changed the document<br>number to<br>KP1/6C/13D/1/TSP/14/0<br>23 on cover page.<br>2.Changed the operating<br>system and processor<br>requirements on the laptop<br>computer.<br>3.Included the requirements<br>of appendix B3, C1,C2 and<br>C3 on the GTPs. | Nancy Wairimu                        | Dr. Eng. Peter<br>Kimemia            |

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 4 of 18</b>  |                                |

## FOREWORD

This specification is prepared by the Standards Department in collaboration with Commercial Services & Sales divisions, both of The Kenya Power and Lighting Company Plc (KPLC) and it lays down requirements for the portable three phase working standard for testing electricity meters. It is intended for procurement of the equipment.

This specification was prepared to establish and promote uniform requirements for portable three-phase watt-hour working standard.

This specification stipulates the minimum requirements for portable three phase working standard for testing electricity meters acceptable for use in the company and it shall be the responsibility of the supplier and manufacturer to ensure that the offered design is of the highest quality and guarantees excellent service to KPLC, good workmanship and good engineering practice in the manufacture of the equipment for KPLC.

The following are members of the team that developed this specification:

| Name           | Division                                 |
|----------------|------------------------------------------|
| Peter Wanyonyi | Commercial Services & Sales              |
| Nancy Wairimu  | Institute of Energy Studies and Research |

**Issued by: Head of Section, Standards Development**

**Signed:**

**Date: 2023-09-28**

**Authorized by: Head of Department, Standards**

**Signed:**

**Date: 2023-09-28**



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 5 of 18</b>  |                                |

### 1. SCOPE

- 1.1. This specification is for a **portable three-phase watt-hour working standard** (further Herein referred simply as "The Standard") for measurement of power and energy in domestic, industrial and commercial meter installations.
- 1.2. This specification covers Class 0.05 **portable three-phase watt-hour working**.
- 1.3. The specification stipulates the minimum requirements of the standard and a laptop, tests and inspection as well as schedule of Guaranteed Technical Particulars.

### 2. NORMATIVE REFERENCES



The following standards contain provisions, which through reference in this text constitute provisions of this specification. For dated editions, the cited edition will apply; for undated editions, the latest edition of the referenced document shall apply.

|                 |                                                                                                                             |
|-----------------|-----------------------------------------------------------------------------------------------------------------------------|
| IEC 61326-1:    | Electrical equipment for measurement, control and laboratory use – EMC requirement – Part 1: General requirements.          |
| IEC 61000-4-30: | Electromagnetic Compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power Quality Measurement methods     |
| IEC 62052-11:   | Electricity Metering Equipment (A.C.) – General Requirements, Tests and Test Conditions – Part 11: Metering Equipment       |
| IEC 62053-21:   | Electricity Metering Equipment – Particular Requirements – Part 21: Static Meters for AC Active Energy (Class 0.5, 1 and 2) |
| IEC 60529       | Degrees of protection provided by Enclosures (IP Code)                                                                      |
| IEC 61010-1:    | Safety requirements for electrical equipment for measurement, control, and laboratory use -Part 1: General requirements     |
| ISO 9001:2015:  | Quality management systems — Requirements                                                                                   |

### 3. DEFINITIONS AND ABBREVIATIONS

For the purpose of this specification the definitions given in the reference standards shall apply and the following abbreviations:

|      |                                                |
|------|------------------------------------------------|
| IEC: | International Electrotechnical Commission      |
| ISO: | International Organization for Standardization |

|                                                                                                    |                                                                                                     |
|----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                 |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                             |

**TITLE:****PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION****Doc. No.****KP1/6C/13D/1/TSP/14/023****Issue No.****5****Revision No.****1****Date of Issue****2023-09-28****Page 6 of 18****4. REQUIREMENTS****4.1. SERVICE CONDITIONS**

The standard shall be suitable for use indoors in tropical areas and harsh climatic conditions including areas exposed to:

- a) At altitudes of up to 2200m above sea level
- b) humidity of up to 95%,
- c) Average ambient temperature of +30°C with a minimum of -1°C and a maximum of +50°C.
- d) Pollution: Design pollution level to be taken as "Heavy" (Pollution level III) for inland and "Very Heavy" (Pollution level IV) for coastal applications in accordance with IEC 60815.
- e) Heavy saline conditions along the coast.

**4.2. DESIGN AND CONSTRUCTION FEATURES**

- 4.2.1. The standard shall be designed and constructed to comply with IEC 61000-4-30 on Power Quality Measurement Methods.
- 4.2.2. The Standard shall be designed to comply with international standard IEC 61326-1 and IEC 61000-4-30 on Electromagnetic compatibility (EMC).
- 4.2.3. The Standard shall be equipped with safety terminals for connection and designed to comply with international safety standard IEC 61010-1.
- 4.2.4. The Standard shall have the capability of testing both single phase and three phase electromechanical and electronic energy meters at site.
- 4.2.5. The standard's accuracy shall be class 0.05.
- 4.2.6. The standard shall be made of hard plastic and /or light aluminium profile or equivalent material to allow a weight of not more than 4kg.
- 4.2.7. The Standard shall be capable of checking meter installation parameters and associated circuits.
- 4.2.8. The Standard shall be used for both direct and via split core transformers (Current clamps) connection in current measurements.
- 4.2.9. The Standard shall measure active, reactive and apparent power and energy measurements for single-phase (2-wire circuit) and three-phase (3-or-4 wire circuits) with integrated error calculator.
- 4.2.10. The standard shall be capable of 4 quadrant power measurement.
- 4.2.11. The standard shall be able to determine the operating burden on instrument transformers for CT and PT.
- 4.2.12. The standard shall be able to perform turns ratio test by simultaneously measurement of both primary and secondary currents in CT connected measuring systems.
- 4.2.13. The standard shall be capable of measuring currents up to a maximum of 2000 A per phase through external clip-on CTs. These Clip-on CTs shall be listed and supplied with equipment.

**Issued by: Head of Section, Standards Development****Authorized by: Head of Department, Standards****Signed:****Signed:****Date: 2023-09-28****Date: 2023-09-28**



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 7 of 18</b>  |                                |

- 4.2.14. The Standard shall have a wide measuring range with auto ranging and definite step ranges on the voltage and current ranges.
- 4.2.15. The Standard shall display vector diagram for analysis of mains conditions and meter connections, phase sequence and relevant instantaneous values.
- 4.2.16. The Standard shall measure and display waveforms.
- 4.2.17. The Standard shall be able to detect circuit connection faults.
- 4.2.18. The Standard shall perform accuracy test, calculate and display the error.
- 4.2.19. The Standard shall be capable of measuring and displaying True RMS values of current and voltage, phase angles, power factor, frequency and phase rotation.
- 4.2.20. Either the standard shall have a display that is at least 6 x 4" with high resolution of at least 320x240 pixels or the display shall be connected wirelessly to a tablet to display the measurement results.
- 4.2.21. The standard shall display voltage sequence indication (U1, U2, and U3).
- 4.2.22. The Standard shall have test data storage of at least 500 measurements, which is capable of being sent, through serial interface to an external PC after measurement by Windows based data management software.
- 4.2.23. The data management software shall be provided without license/ license costs
- 4.2.24. The standard shall measure harmonics in voltage and current up to 40<sup>th</sup> THD.
- 4.2.25. The Standard shall have pulsed output for energy (galvanic isolation).
- 4.2.26. The Standard shall have features like Hold-function to read the instantaneous values. During meter testing and integrated START/STOP-push button for a quick visual meter check, without scanning head.
- 4.2.27. The standard shall have a provision for insertion of a scanning head for testing both electromechanical and static (electronic) meters.
- 4.2.28. The standard shall have selective power measurement.
- 4.2.29. The standard shall be able to test power and energy registers.
- 4.2.30. The standard shall be delivered with a MS-Windows analyzer and simulation software for vectorial diagrams.
- 4.2.31. The standard shall be fitted with a handle and a belt to make it possible for the operator to be able to hang it on shoulder or neck while operating at site.
- 4.2.32. The Photoelectric Scanning head shall be suitable for use with both LED pulsed electronic meters and Ferraris meters, selectable via a selector switch.
- 4.2.33. The Photoelectric Scanning head support shall have a spring mechanism with different adjusting possibilities to make it possible to attach as desired the scanning head to the meter.
- 4.2.34. The standard shall be fitted with a battery pack to power the standard to enable testing without the aid of an auxiliary power source.
- 4.2.35. The standard shall be powered through either of the following conditions; an auxiliary source, a battery pack or grid supply.

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 8 of 18</b>  |                                |

4.2.36. The Standard shall have a multirange power supply to accommodate inputs of up to 265 volts.

4.2.37. The standard shall have overvoltage protection.

#### 4.3. RATING

4.3.1. The Standard shall comply with the ratings outlined below:

| NO. | ITEM DESCRIPTIONS                                        | REQUIREMENT                                                                                                  |
|-----|----------------------------------------------------------|--------------------------------------------------------------------------------------------------------------|
|     |                                                          | Class 0.05                                                                                                   |
| 1.  | Auxiliary Supply                                         | 80.....265VAC , 47.....63 Hz                                                                                 |
| 2.  | Power Supply through test voltage                        | 80.....265VAC , 47.....63 Hz                                                                                 |
| 3.  | Power Consumption                                        | ≤ 138 VA                                                                                                     |
| 4.  | Voltage Measurement                                      | 1mV ..... 300V                                                                                               |
| 5.  | Influence of auxiliary voltage on the measuring circuits | ≤ ± 0.005 % at 10 % variation                                                                                |
| 6.  | Test Current, direct connection                          | 1 mA.....12 A                                                                                                |
| 7.  | Test Current, clamp-on                                   | 5 mA.....120 A                                                                                               |
| 8.  | Fundamental Frequency                                    | 15.....70 Hz                                                                                                 |
| 9.  | Measuring modes                                          | 4 Wire active/-reactive/apparent<br>3 Wire active /- reactive/apparent<br>2 Wire active /- reactive/apparent |
| 10. | Accuracy class                                           | 0.05                                                                                                         |
| 11. | Voltage measurement error (90.....300 V)                 | ≤ 0.025 %                                                                                                    |
| 12. | Current measurement error, direct (1 mA.....12 A)        | ≤ 0.025 %                                                                                                    |
| 13. | Current measurement error, via CT(500 mA.....120 A)      | ≤ 0.1 %                                                                                                      |
| 14. | Current measurement error, via CT ( up to 2000 A)        | ≤ 0.2 %                                                                                                      |
| 15. | Power/energy measurement error                           | ≤ 0.05 % (direct measurement)                                                                                |
| 16. | Power/energy measurement error                           | ≤ 0.05 % ( with CT's)                                                                                        |
| 17. | Angle measurement (direct measurement)                   | ≤ ± 0.01°                                                                                                    |
| 18. | Angle measurement (with clamps/CTs)                      | ≤ ± 0.1°                                                                                                     |
| 19. | Frequency measurement                                    | ± 0.01 Hz                                                                                                    |
| 20. | Temperature range (Operation)                            | -1.....+ 50 °C                                                                                               |
| 21. | Relative Humidity                                        | Up to 80%Not condensing                                                                                      |

**Issued by: Head of Section, Standards Development**

**Authorized by: Head of Department, Standards**

**Signed:**

**Signed:**

**Date: 2023-09-28**

**Date: 2023-09-28**



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

**Doc. No.** KP1/6C/13D/1/TSP/14/023  
**Issue No.** 5  
**Revision No.** 1  
**Date of Issue** 2023-09-28  
**Page 9 of 18**

| NO. | ITEM DESCRIPTIONS    | REQUIREMENT                           |
|-----|----------------------|---------------------------------------|
|     |                      | Class 0.05                            |
| 22. | Sampling Rate        | Minimum 16 bit 504 samples per period |
| 23. | Safety protection    | IEC 61010-1                           |
| 24. | Degree of protection | Minimum IP-30                         |
| 25. | Certification        | CE, IEC 61000-4-30.                   |

#### 4.4. COMPUTER REQUIREMENTS

4.4.1. The Standard shall be supplied with one laptop computer loaded with the operating software.

| Description      | Mandatory Minimum Requirements                                                                                                                                                              |
|------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Processor        | Intel Core i7 11 <sup>th</sup> Generation or later (2.60GHz 1600MHz 3MB, 8 Cores)                                                                                                           |
| RAM              | 8GB DDR4-2133MHz SODIMM                                                                                                                                                                     |
| Operating System | Windows 11 pro 64 bit                                                                                                                                                                       |
| Optical Drive    | Super Multi DVD-RW or DVD Recordable Burner                                                                                                                                                 |
| Hard Disk        | 1TB 7200 rpm Hard Drive or 500 SSD                                                                                                                                                          |
| Display Panel    | 15.6" FHD LED Glossy (1920x1080) with integrated Webcam 720p camera                                                                                                                         |
| Graphics         | Integrated Intel HD Graphics 520                                                                                                                                                            |
| Internal Audio   | Integrated HD audio internal speaker (standard) or Stereo with Dolby Audio TM, 1xMic Headphones Combo                                                                                       |
| Communications   | 56K Modem, Integrated Intel Gigabit Network Connection (10/100/1000 NIC)                                                                                                                    |
| Wireless         | Intel 802.12 AC WLAN and Bluetooth(R)                                                                                                                                                       |
| Interfaces       | VGA, MDP, 4-in-1 Card Reader, Smart Card Reader. RJ-45, Headphone and Microphone Jack, Mechanical Docking, 2 x USB 3.0, W/WAN SIM, Express Card Slot, 1 HDMI port, Bluetooth, Wi-Fi enabled |
| Pointing Devices | Touchpad with scroll zone, Two Pick Buttons or Pick Stick, Two Pick Buttons                                                                                                                 |
| Keyboard         | Keyboard with Number Pad – English (Standard)                                                                                                                                               |
| Mouse            | External USB Mouse                                                                                                                                                                          |
| Warranty         | 1 Year                                                                                                                                                                                      |
| Power            | 4-cell 41WHr Lithium-ion Battery; External AC adapter                                                                                                                                       |
| Power Supply     | 230V AC, 50 Hz, British plugs                                                                                                                                                               |
| Carrying Case    | Genuine Leather Carrying Case                                                                                                                                                               |

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 10 of 18</b> |                                |

**5. ACCESSORIES**

The standard shall have

- A set of Test leads (quick connecting cable set),
- CT's (120 A), Accessories,
- Photoelectric Scanning head with its support contained, and
- An additional Scanning head with magnet and
- Sucker for expedient testing of LEDs.

**6. TESTS REQUIREMENTS**

The Standard shall be inspected and tested in accordance with the requirements of IEC standards and other relevant standards and provisions of this specification.

**7. MARKING, LABELLING AND PACKING**

7.1. The Standard shall be marked legibly and indelibly in English and with at least 4mm figure height with the following information:

- Name or trade mark of the manufacturer;
- Country of origin;
- Type/model;
- Serial no;
- Nominal input voltage and frequency
- Power Consumption in Watts or VA's
- Fuse ratings;
- The inscription "Property of KPLC";
- Year of manufacture.

7.2. The Standard shall be packaged in such a manner as to minimize damage and entry of moisture during transportation, storage and handling.

7.3. A rugged transportation case shall be provided to store and/or transport the standard and its accessories.

**Issued by: Head of Section, Standards Development**

**Signed:**

**Date: 2023-09-28**

**Authorized by: Head of Department, Standards**

**Signed:**

**Date: 2023-09-28**



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 11 of 18</b> |                                |



**APPENDICES**

**A. TESTS AND INSPECTION (Normative)**

- A.1 It shall be the responsibility of the manufacturer to perform or to have performed all the tests specified.
- A.2 Copies of Type Test Certificates and Type Test Reports issued by a third-party testing laboratory that is accredited to ISO/IEC 17025 shall be submitted with the tender for the purpose of technical evaluation. A copy of the accreditation certificate to ISO/IEC 17025 for the testing laboratory shall also be submitted. Any translations of certificates and test reports into English language shall be signed and stamped by the Testing Laboratory that carried out the tests.
- A.3 The Standard shall be subject to acceptance tests at the manufacturer's premises before dispatch. Acceptance tests shall be witnessed by two Engineers appointed by The Kenya Power and Lighting Company Plc. (KPLC).

**A.4 Testing Facility**

- A.4.1 The bidder shall provide current e-mail address, fax and telephone numbers and contact person at the Testing Laboratory where Type Tests were carried out.
- A.4.2 All test and measuring equipment to be used during acceptance testing shall have been calibrated and copies of valid calibration certificates shall be provided to KPLC Engineers. A detailed list of workshop tools, test/measuring equipment and list of tests that can be carried out by the manufacturer shall be submitted with the tender for evaluation.
- A.5 Test reports for the Standard shall be submitted to The Kenya Power and Lighting Company for approval before shipment.
- A.6 During delivery of Standard, KPLC will inspect them and may perform or have performed any of the relevant tests in order to verify compliance with the specification. The supplier shall replace/rectify without charge to KPLC, the Standard that fails to meet any or all of the requirements in the specification.

|                                                                                                    |                                                                                                     |
|----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                 |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**


|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 12 of 18</b> |                                |

**B. QUALITY MANAGEMENT SYSTEM (Normative)**

- B.1 The bidder shall submit a quality assurance plan (QAP) that will be used to ensure that the Standard design, material, workmanship, tests, service capability, maintenance and documentation, will fulfil the requirements stated in the contract documents, standards, specifications and regulations. The QAP shall be based on and include relevant parts to fulfil the requirements of ISO 9001:2015.
- B.2 The Manufacturer's Declaration of Conformity to applicable standards, this specification and copies of quality management certifications including copy of valid and relevant ISO 9001 certificate shall be submitted with the tender for evaluation.
- B.3 The bidder shall indicate the delivery time of the Standard, experience, a detailed list and contact addresses (including e-mail) of previous utilities to which the standard has been supplied. The number of standards sold over a period of 5 years shall not be less than 100 standards.

**C. DOCUMENTATION AND DEMONSTRATION (Normative)**

- C.1 The bidder shall submit its tender complete with technical documents required by Appendix D (Guaranteed Technical Particulars) for tender evaluation. The documents to be submitted (all in English language) for tender evaluation shall include the following:
- Fully filled clause by clause Guaranteed Technical Particulars (GTP) signed by the manufacturer,
  - Copies of the manufacturer's catalogues, brochures, Standard drawings and wiring diagrams and technical data showing description leaflet, programming details and manuals,
  - Sales records for the last five years and at least four customer reference letters,
  - Details of manufacturing capacity and the manufacturer's experience.
  - Copies of required type test certificates and type test reports by a third-party testing laboratory accredited to ISO/IEC 17025,
  - Copy of accreditation certificate to ISO/IEC 17025 for the third-party testing laboratory,
  - Manufacturer is guaranteed and guarantee; subject to 60 months from date of acceptance by KPLC laboratory.
  - Manufacturer's letter of authorization, copy of the manufacturer's ISO 9001:2015 certificate, ISO 17025: 2017 certificate.


|                                                                                                    |                                                                                                     |
|----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                 |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 13 of 18</b> |                                |

- C.2 The successful bidder (supplier) shall submit the following documents/details to The Kenya Power & Lighting Company for approval before manufacture:
- a) Fully filled clause by clause Guaranteed Technical Particulars (GTP) signed by the manufacturer,
  - b) Design drawings and wiring diagrams of the Standard,
  - c) Original software, software manuals and operation manuals shall be submitted,
  - d) Quality assurance plan (QAP) that will be used to ensure that the design, material, workmanship, tests, service capability, maintenance and documentation will fulfill the requirements stated in the contract documents, standards, specifications and regulations. The QAP shall be based on and include relevant parts to fulfill the requirements of ISO 9001:2015,
  - e) Detailed test program to be used during factory testing,
  - f) Marking details and method to be used in marking the Standard,
  - g) Manufacturer's undertaking to ensure adequacy of the design, good engineering practice, adherence to the specification and applicable standards and regulations as well as ensuring good workmanship in the manufacture of the Standard for The Kenya Power & Lighting Company,
  - h) Packaging details (including packaging materials).
- C.3 The successful bidder and Manufacturer shall demonstrate the design, connection and measurement of all parameters of all the parameters the standard is able to measure at their cost to at least 20 KPLC staff locally.

|                                                                                                    |                                                                                                     |
|----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b>                                           | <b>Authorized by: Head of Department, Standards</b>                                                 |
| <b>Signed:</b>  | <b>Signed:</b>  |
| <b>Date: 2023-09-28</b>                                                                            | <b>Date: 2023-09-28</b>                                                                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 14 of 18</b> |                                |

**D. GUARANTEED TECHNICAL PARTICULARS (Normative)**

*To be filled and signed by the manufacturer and submitted together with relevant copies of the Manufacturer's catalogues, brochures, drawings, technical data, sales records for previous five years, four customer reference letters, details of suppliers' capacity and experience; and copies of complete type test certificates and test reports for tender evaluation, all in English Language)*

**Tender No.** .....

**Bidder's name and Address**.....

| Clause number | KPLC requirement                                                                                                                                                                       | Bidder's offer |
|---------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
|               | Manufacturer's Name and address                                                                                                                                                        | State          |
|               | Country of Manufacture                                                                                                                                                                 | State          |
|               | Bidder's Name and address                                                                                                                                                              | State          |
| 1.            | Scope                                                                                                                                                                                  | State          |
| 2.            | Normative references                                                                                                                                                                   | State          |
| 3.            | Definitions and Abbreviations                                                                                                                                                          | State          |
| 4             | <b>Requirements</b>                                                                                                                                                                    |                |
| 4.1           | <b>Service Conditions</b>                                                                                                                                                              |                |
|               | The Standard shall be suitable for use indoors in tropical areas and harsh climatic conditions including areas exposed to:                                                             |                |
| a)            | At altitudes of up to 2200m above sea level and,                                                                                                                                       | State          |
| b)            | humidity of up to 95%                                                                                                                                                                  | State          |
| c)            | Average ambient temperature of +30°C with a minimum of -1°C and a maximum of +50°C.                                                                                                    | State          |
| d)            | Pollution: Design pollution level to be taken as "Heavy" (Pollution level III) for inland and "Very Heavy" (Pollution level IV) for coastal applications in accordance with IEC 60815. | State          |
| e)            | Heavy saline conditions along the coast                                                                                                                                                | State          |
| 4.2.          | <b>DESIGN FEATURES</b>                                                                                                                                                                 |                |
| 4.2.1         | The Standard of manufacture                                                                                                                                                            | State          |
| 4.2.2         | The Standard shall be designed to comply with international standard EN 61326-1 and EN 61000-4-30 on Electromagnetic compatibility (EMC)                                               | State          |
| 4.2.3         | The Standard shall be equipped with safety terminals for connection to prevent shock and designed to comply with international safety standard IEC 61010-1.                            | State          |

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 15 of 18</b> |                                |

| Clause number | KPLC requirement                                                                                                                                              | Bidder's offer |
|---------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
| 4.2.4         | The Standard the capability of testing both single phase and three phase electromechanical and electronic energy meters at site                               | State          |
| 4.2.5         | The Standard's accuracy                                                                                                                                       | State          |
| 4.2.6         | The Standard material of manufacture and weight.                                                                                                              | State          |
| 4.2.7         | The Standard capability of checking meter installation parameters and associated circuits.                                                                    | State          |
| 4.2.8         | The Standard shall use for both direct and via split core transformers (Current clamps) connection in current measurements.                                   | State          |
| 4.2.9         | The standard's measurands                                                                                                                                     | State          |
|               | Capability of measuring single-phase (2-wire circuit) and three-phase (3-or-4 wire circuits) with integrated error calculator                                 | State          |
| 4.2.10        | The standard shall be capable of 4 quadrant power measurement                                                                                                 | State          |
| 4.2.11        | The standard shall be able to determine the operating burden on instrument transformers for CT and PT                                                         | State          |
| 4.2.12        | The standard shall be able to perform turns ratio test by simultaneously measurement of both primary and secondary currents in CT connected measuring systems | State          |
| 4.2.13        | The standard shall be capable of measuring currents up to a maximum of 2000 A per phase through external clip-on CTs.                                         | State          |
|               | Clip-on CTs shall be listed and supplied with equipment                                                                                                       | State          |
| 4.2.14        | The Standard shall have a wide measuring range with auto ranging and definite step ranges on the voltage and current ranges.                                  | State          |
| 4.2.15        | The Standard shall display vector diagram for analysis of mains conditions and meter connections and relevant instantaneous values                            | State          |
| 4.2.16        | The Standard shall measure and display waveforms                                                                                                              | State          |
| 4.2.17        | The Standard shall be able to detect circuit connection faults                                                                                                | State          |
| 4.2.18        | The Standard shall perform accuracy test, calculate and display the error                                                                                     | State          |
| 4.2.19        | The Standard shall be capable of measuring and displaying True RMS values of current and voltage, phase angles, power factor, frequency and phase rotation    | State          |
| 4.2.20        | The display shall be of at least 6 x 4" with high resolution of at least 320x240 pixels                                                                       | State          |
|               | The display shall be connected wirelessly to a tablet to display the measurement results.                                                                     | State          |
| 4.2.21        | The standard shall display voltage sequence indication (U1, U2, and U3)                                                                                       | State          |
| 4.2.22        | The standard storage capacity                                                                                                                                 | State          |

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

**Doc. No.** KP1/6C/13D/1/TSP/14/023  
**Issue No.** 5  
**Revision No.** 1  
**Date of Issue** 2023-09-28  
**Page 16 of 18**

| Clause number | KPLC requirement                                                                                                                                                                      | Bidder's offer |
|---------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
| 4.2.23        | Data management software shall be provided without license/ license cost                                                                                                              | State          |
| 4.2.24        | The standard shall measure harmonics in voltage and current up to 40 <sup>th</sup> THD                                                                                                | State          |
| 4.2.25        | The Standard shall have pulsed output for energy (galvanic isolation)                                                                                                                 | State          |
| 4.2.26        | The Standard features                                                                                                                                                                 | State          |
| 4.2.27        | The standard shall have a provision for insertion of a scanning head for testing both electromechanical and static (electronic) meters                                                | State          |
| 4.2.28        | The standard shall have selective power measurement                                                                                                                                   | State          |
| 4.2.29        | The standard shall be able to test power and energy registers                                                                                                                         | State          |
| 4.2.30        | The standard shall be delivered with a MS-Windows analyzer and simulation software for vectorial diagrams                                                                             | State          |
| 4.2.31        | The standard shall be fitted with a belt to make it possible for the operator to be able to hang it on shoulder or neck while operating at site.                                      | State          |
| 4.2.32        | The Photoelectric Scanning head shall be suitable for use with both LED pulsed electronic meters and Ferraris meters, selectable via a selector switch.                               | State          |
| 4.2.33        | The Photoelectric Scanning head support shall have a spring mechanism with different adjusting possibilities to make it possible to attach as desired the scanning head to the meter. | State          |
| 4.2.3.4       | The standard power supply                                                                                                                                                             | State          |
| 4.2.3.5       | The standard powering conditions                                                                                                                                                      | State          |
| 4.2.3.6       | The Standard shall have a multirange power supply to accommodate inputs of up to 265 volts.                                                                                           | State          |
| 4.2.3.7       | Overvoltage protection                                                                                                                                                                | State          |
| <b>4.3</b>    | <b>Rating</b>                                                                                                                                                                         |                |
|               | Auxiliary Supply                                                                                                                                                                      | State          |
|               | Power Supply through test voltage                                                                                                                                                     | State          |
|               | Power Consumption                                                                                                                                                                     | State          |
|               | Voltage Measurement                                                                                                                                                                   | State          |
|               | Influence of auxiliary voltage on the measuring circuits                                                                                                                              | State          |
|               | Test Current, direct connection                                                                                                                                                       | State          |
|               | Test Current, clamp-on                                                                                                                                                                | State          |
|               | Fundamental Frequency                                                                                                                                                                 | State          |
|               | Measuring modes                                                                                                                                                                       | State          |
|               | Accuracy class                                                                                                                                                                        | State          |
|               | Voltage measurement error (90.....300 V)                                                                                                                                              | State          |
|               | Current measurement error, direct (1 mA.....12 A)                                                                                                                                     | State          |

**Issued by: Head of Section, Standards Development**

**Authorized by: Head of Department, Standards**

**Signed:**

**Signed:**

**Date: 2023-09-28**

**Date: 2023-09-28**



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

**Doc. No.** KP1/6C/13D/1/TSP/14/023  
**Issue No.** 5  
**Revision No.** 1  
**Date of Issue** 2023-09-28  
**Page 17 of 18**

| Clause number | KPLC requirement                                                                                                                                            | Bidder's offer |
|---------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
|               | Current measurement error, via CT(500 mA.....120 A)                                                                                                         | State          |
|               | Current measurement error, via CT ( up to 2000 A)                                                                                                           | State          |
|               | Power/energy measurement error                                                                                                                              | State          |
|               | Power/energy measurement error                                                                                                                              | State          |
|               | Angle measurement (direct measurement)                                                                                                                      | State          |
|               | Angle measurement (with clamps/CTs)                                                                                                                         | State          |
|               | Frequency measurement                                                                                                                                       | State          |
|               | Temperature range (Operation)                                                                                                                               | State          |
|               | Relative Humidity                                                                                                                                           | State          |
|               | Sampling Rate                                                                                                                                               | State          |
|               | Safety protection                                                                                                                                           | State          |
|               | Degree of protection                                                                                                                                        | State          |
|               | Certification                                                                                                                                               |                |
| <b>4.4.</b>   | <b>COMPUTER REQUIREMENTS</b>                                                                                                                                |                |
|               | The Standard shall be supplied with one laptop computer loaded with the operating software and the specifications outlined as per the table in clause 4.4.1 | Specify        |
| <b>5.</b>     | <b>ACCESSORIES</b>                                                                                                                                          |                |
|               | The accessories associated with the Standard                                                                                                                | List           |
| <b>6.</b>     | <b>TESTS REQUIREMENT</b>                                                                                                                                    | State          |
| <b>7</b>      | <b>MARKING, LABELLING AND PACKING</b>                                                                                                                       |                |
| 7.1           | The Standard shall be marked legibly and indelibly in English with the following information:                                                               | State          |
| a)            | Name or trade mark of the manufacturer;                                                                                                                     | State          |
| b)            | Country of origin;                                                                                                                                          | State          |
| c)            | Type/model;                                                                                                                                                 | State          |
| d)            | Serial no:                                                                                                                                                  | State          |
| e)            | Nominal input voltage and frequency                                                                                                                         | State          |
| f)            | Power Consumption in Watts or VA's                                                                                                                          | State          |
| g)            | Fuse ratings                                                                                                                                                | State          |
| i)            | The inscription "Property of KPLC"                                                                                                                          | State          |
| j)            | Year of manufacture.                                                                                                                                        | State          |
| 7.2           | The Standard shall be packaged in such a manner as to minimize damage and entry of moisture during transportation and handling.                             | State          |

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| <b>Signed:</b>                                           | <b>Signed:</b>                                      |
| <b>Date: 2023-09-28</b>                                  | <b>Date: 2023-09-28</b>                             |



**TITLE:**  
**PORTABLE THREE PHASE  
WORKING STANDARD FOR  
TESTING ELECTRICITY  
METERS - SPECIFICATION**

|                      |                                |
|----------------------|--------------------------------|
| <b>Doc. No.</b>      | <b>KP1/6C/13D/1/TSP/14/023</b> |
| <b>Issue No.</b>     | <b>5</b>                       |
| <b>Revision No.</b>  | <b>1</b>                       |
| <b>Date of Issue</b> | <b>2023-09-28</b>              |
| <b>Page 18 of 18</b> |                                |

| Clause number | KPLC requirement                                                                                           | Bidder's offer   |
|---------------|------------------------------------------------------------------------------------------------------------|------------------|
| 7.3           | A rugged transportation case shall be provided to store and/or transport the standard and its accessories. | State            |
|               | <b>APPENDICES</b>                                                                                          |                  |
| <b>A.</b>     | <b>TESTS AND INSPECTION (Normative)</b>                                                                    |                  |
| A.1           | Responsibility of carrying out tests                                                                       | State            |
| A.2           | Copies of Type Test Reports to be submitted with tender                                                    | Provide          |
| A.3           | Contacts of testing laboratory                                                                             | Provide          |
| A.4           | Acceptance tests at the manufacturers works                                                                | State compliance |
| A.5           | Test certificates to be submitted by supplier to KPLC for approval before supply/delivery                  | State compliance |
| A.6           | Inspection at the stores and replacement of rejected items                                                 | State compliance |
| <b>B</b>      | <b>Quality Management System</b>                                                                           |                  |
| B.1           | Quality Assurance Plan                                                                                     | Provide          |
| B.2           | Copy of ISO 9001:2015 Certificate                                                                          | Provide          |
| B.3           | Manufacturer's experience                                                                                  | State            |
|               | Manufacturing Capacity (units per month)                                                                   | State            |
|               | List of previous customers                                                                                 | State            |
|               | Customer reference letters                                                                                 | State            |
| <b>C</b>      | <b>Documentation</b>                                                                                       |                  |
| C.1           | Documents submitted with tender                                                                            | State compliance |
| C.2           | Documents to be submitted by supplier to KPLC for approval before manufacture                              | State compliance |
| C.3           | Demonstration locally                                                                                      | State            |
|               | Statement of compliance to specification (indicate deviations if any & supporting documents)               | State compliance |

**Words like "confirmed", "noted", "agreed" "As per KPLC specification" etc. shall not be accepted and shall be considered as non-responsive where a specific response is called for.**

.....  
**Manufacturer's Name, Signature, Stamp and Date**

|                                                          |                                                     |
|----------------------------------------------------------|-----------------------------------------------------|
| <b>Issued by: Head of Section, Standards Development</b> | <b>Authorized by: Head of Department, Standards</b> |
| Signed:                                                  | Signed:                                             |
| Date: 2023-09-28                                         | Date: 2023-09-28                                    |

